

HERITAGE  
MEDICAL

Analizator składu ciała ACCUNIQ

Deklaracja zgodności i certyfikaty jakości

# EC Declaration of Conformity

*Manufacturer:*

SELVAS Healthcare, Inc.  
29, Gongdan 4-ro, Jillyang-eup, Gyeongsan-si, Gyeongsangbuk-do, 38470 Republic of Korea

We, the manufacturer, herewith declare that the products

**Body Composition Analyzer**

**Model: ACCUNIQ BC310**

*UMDNS-Code: 17417*

meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to class IIa (body composition analyzer) according to Annex IX of the Directive 93/42/EEC. It bears the mark

**CE 0197**

The product concerned has been manufactured under a quality management system according to Annex V of Directive 93/42/EEC, and the essential requirement of Annex I pertaining to medical devices

Compliance of the designated product with the Directive 93/42/EEC has been assessed and certified by the Notified Body

**TÜV Rheinland LGA Products GmbH**

**Tillystraße 2**

**90431 Nürnberg**

**Country : Germany**

Certificate No.: DD 60105723 0001

Expiry date: 25.11.2020

following the procedure relating to the EC Declaration of Conformity set out in Annex V of Directive 93/42/EEC.

This Declaration of conformity is valid in connection with the release document for the respective batch of produced devices

The above mentioned declaration of conformity is exclusively under the responsibility of

Company: SELVAS Healthcare, Inc.

Address: 29, Gongdan 4-ro, Jillyang-eup, Gyeongsan-si, Gyeongsangbuk-do, 38470 Republic of Korea

Factory: 29, Gongdan 4-ro, Jillyang-eup, Gyeongsan-si, Gyeongsangbuk-do, 38470 Republic of Korea

Gyeongsan-si, 1, SEP 2016

Place, date

Jin-Kyu Baek/QMR

Legally binding signature, Function



**EC Certificate**  
**Directive 93/42/EEC Annex V**  
**Production Quality Assurance**  
**Medical Devices**

**Registration No.:** DD 60114162 0001

**Report No.:** 12022725 003

**Manufacturer:** SELVAS Healthcare, Inc.  
29, Gongdan 4-ro, Jillyang-eup  
Gyeongsan-si, Gyeongsangbuk-do, 38470  
Republic of Korea

**Products:**

- Non-Invasive Blood Pressure Monitors
- Body Composition Analyzers

**Expiry Date:** 2020-11-25

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

**Effective Date:** 2016-10-06

**Date:** 2016-10-06




Notified Body

*M. Aihara*  
M.Sc. M. Aihara

**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

**Business Stream Products  
Certification Department**



TÜV Rheinland LGA Products GmbH · 90431 Nürnberg

SELVAS Healthcare, Inc.  
Mr. J. K. Baek  
29, Gongdan 4-ro, Jillyang-eup  
GYEONGSAN-SI, GYEONGSANGBUK-DO, 384  
REPUBLIC OF KOREA

**Contact**

Tel. +49 911 655-5225  
Mail [service@de.tuv.com](mailto:service@de.tuv.com)

Date October 06, 2016

**Application for** : QMS Produktion, Anhang V MDD  
Certificate No. : DD 60114162 Sheet 0001  
Device : Only for QM-System audit  
Test requirement : Richtlinie 93/42/EWG

Dear Mr. Baek,

The Extraordinary audit of your quality management system to cover company name change was performed.

The audit team confirmed that your quality management system is applied effectively with respect to the above mentioned requirements.

Enclosed please find the new certificate No. DD 60114162 0001 replacing the previous certificate DD 60105723 0001.

With issue date of the new certificate, the previous certificate has become invalid.

Kind regards

Certification body



M.Sc. M. Aihara

Test sample: no, documentation available

TÜV Rheinland  
LGA Products GmbH

Tillystraße 2  
90431 Nürnberg

Tel. +49 911 655-5225  
Fax +49 911 655-5226  
Mail [service@de.tuv.com](mailto:service@de.tuv.com)  
Web [www.tuv.com/safety](http://www.tuv.com/safety)

Board of Management

Dipl.-Ing.  
Jörg Mähler, Spokesman

Dipl.-Kfm.  
Dr. Jörg Schlösser

Chairman of the  
Supervisory Board

Dipl.-Ing.  
Raif Scheller

Nuremberg HRB 26013  
VAT No.: DE 811835490



# Certificate

The Certification Body of  
TÜV Rheinland LGA Products GmbH

hereby certifies that the organization  
**SELVAS Healthcare, Inc.**  
**155, Shinseong-ro, Yuseong-gu**  
**Daejeon 34109**  
**Republic of Korea**

has established and applies a quality management system for medical devices  
for the following scope:

**Manufacture and Distribution of Non-Invasive Blood Pressure  
Monitors and Body Composition Analyzers**

Proof has been furnished that the requirements specified in

**EN ISO 13485:2016**

are fulfilled. The quality management system is subject to yearly surveillance.

Effective Date: 2018-11-26  
Certificate Registration No.: SX 60134339 0001  
An audit was performed. Report No.: 12022725 007  
This Certificate is valid until: 2021-05-20

Certification Body



Date 2018-11-26



*M. Aihara*

M.Sc. M. Aihara

**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**  
Tel +49 221 806-1371 Fax +49 221 806-3935 e-mail cert-validity@de.tuv.com http://www.tuv.com/safety

**Business Stream Products  
Certification Department**



Precisely Right.

TUV Rheinland LGA Products GmbH · 90431 Nürnberg

SELVAS Healthcare, Inc.  
Mr. J. K. Baek  
155, Shinseong-ro, Yuseong-gu  
DAEJEON 34109  
REPUBLIC OF KOREA

Contact

Tel. +49 911 655-5225  
Mail [service@de.tuv.com](mailto:service@de.tuv.com)

Date November 26, 2018

**Application for** : QMS  
Certificate No. : SX 60134339 Sheet 0001  
Device : Only for QM-System audit  
Test requirement : EN ISO 13485:2016

Dear Mr. Baek,

Your Quality Management System has been tested and found to be in accordance with the above mentioned requirements.

Enclosed please find the certificate  
No. SX 60134339 0001.

Kind regards

Certification body

A handwritten signature in cursive script, appearing to read 'M. Aihara'.

M.Sc. M. Aihara

Test sample: no, documentation available

TÜV Rheinland  
LGA Products GmbH

Tillystraße 2  
90431 Nürnberg

Tel. +49 911 655-5225  
Fax +49 911 655-5226  
Mail [service@de.tuv.com](mailto:service@de.tuv.com)  
Web [www.tuv.com/safety](http://www.tuv.com/safety)

Board of Management

Dipl.-Ing.  
Jörg Mähler, Spokesman

Dipl.-Kfm  
Dr. Jörg Schösser

Chairman of the  
Supervisory Board

Dipl.-Ing.  
Ralf Scheller

Nuremberg HRB 26013  
VAT No.: DE 811835490



HERITAGE  
MEDICAL

Więcej informacji na temat analizatora ACCUINIQ znajdą na naszej stronie:

[urzadzenia-medyczne.com.pl](http://urzadzenia-medyczne.com.pl)

Zapraszamy również do kontaktu telefonicznego lub mailowego:

+48 660 983 574

+48 12 263 73 27

+48 793 930 812

[biuro@heritage-medical.pl](mailto:biuro@heritage-medical.pl)